EPA Region 7 Laboratory Analysis Information

04/05/2019 08:54

ASR: 8209

Analysis Name: Metals in Solids by ICP-AES

Analysis Request No 1

Analysis Short Name Met S.3G Parameter Class: Metals

Matrix: Solid Analysis Status: Current

Analysis Summary:

This analysis is performed following R7 Method 3122.3 and provides analytical data for up to 23 elements in solids. Many of the elements in soil are naturally occurring, but some elements may have elevated concentrations due to improper waste disposal or pesticide applications that contained metals such as arsenic. Other types of solid samples that can be analyzed include dust, sediment, industrial waste, sludges, etc. Solid samples analyzed for metals originate from an array of different sources, as well as different programs, such as Superfund and RCRA. However, this method still satisfies all of the applicable program requirements for both Superfund and RCRA (and potentially other programs as well).

The presence and concentrations of these elements are determined by the soil sample going through digestion phase followed by analysis of the extract. An aliquot, 1 g, of the sample is digested by adding 1 milliliters nitric acid (HNO3) and 1.5 milliliters of hydrochloric acid (HCI) and then heated in a block digestor at 85-95 C. The digested sample is brought up to 50 milliliters with DI water and then an aliquot is centrifuged for analysis. The sample is analyzed by inductively coupled argon plasma (ICAP) atomic emission spectroscopy. The detection limits range from 0.101 ppm (mg/kg) for beryllium to 23.4 ppm for sodium, and are typically in the low (less than 2) ppm range. Matrix interferences sometimes occur and may complicate the analytical process. If there is a matrix interference, dilutions may be performed and higher detection limits reported. Potential other sources of interferences include contaminated laboratory equipment and/or chemicals. However, analytical interferences are typically the result of the sample matrix and not laboratory materials.

Method: RLAB Method 3122.3G Date Adopted: 04/06/2018 Date Replaced:

Method Desc:

EPA Region 7 RLAB Method 3122.3G

Method Title:

Analysis of Metals by Inductibely Coupled Plasma - Atomic Emission Spectrometry

Method Summary:

(Not Available)

Base MethodTitleTypeSW846 6010BAnalysis of Metals by ICPAnalysis

Capable Labs: EPA (In-House)

ESAT (In-House)
CLP (Out-Source)
REST (Out-Source)

Previous Method: RLAB Method 3122.3F Date Adopted: 09/23/2013 Date Replaced: 04/06/2018

Sample Holding Time 180 Days Extract Holding Time: 0 Days No Of Containers: 1

Container Type: 8 oz glass Weight Type: Dry No Of Tags: 1

Preservative: 4 Deg C

ASR: 8209

Analysis Name: Metals in Solids by ICP-AES

Analysis Request No 1

Sampling Info: (Not Available)

Sampling Narative:

(Not Available)

	Report Flag		Analyte Name	CAS Number	TRL	Con Of	Units
Default	Requested	EPA				Concern	
Yes	Yes	Yes	Aluminum	7429-90-5	5		mg/kg
Yes	Yes	Yes	Antimony	7440-36-0	2		mg/kg
Yes	Yes	Yes	Arsenic	7440-38-2	5		mg/kg
Yes	Yes	Yes	Barium	7440-39-3	2		mg/kg
Yes	Yes	Yes	Beryllium	7440-41-7	1		mg/kg
Con	No	No	Bismuth	7440-69-9	5		mg/kg
Con	No	No	Boron	7440-42-8			mg/kg
Yes	Yes	Yes	Cadmium	7440-43-9	1		mg/kg
Yes	Yes	Yes	Calcium	7440-70-2	50		mg/kg
Yes	Yes	Yes	Chromium	7440-47-3	2		mg/kg
Yes	Yes	Yes	Cobalt	7440-48-4	1		mg/kg
Yes	Yes	Yes	Copper	7440-50-8	1		mg/kg
Yes	Yes	Yes	Iron	7439-89-6	5		mg/kg
Yes	Yes	Yes	Lead	7439-92-1	5		mg/kg
Non	No	No	Lithium	7439-93-2	2		mg/kg
Yes	Yes	Yes	Magnesium	7439-95-4	50		mg/kg
Yes	Yes	Yes	Manganese	7439-96-5	5		mg/kg
Yes	Yes	Yes	Molybdenum	7439-98-7	2		mg/kg
Yes	Yes	Yes	Nickel	7440-02-0	2		mg/kg
Yes	Yes	Yes	Potassium	7440-09-7	50		mg/kg
Yes	Yes	Yes	Selenium	7782-49-2	10		mg/kg
Con	No	No	Silicon	7440-21-3			mg/kg
Yes	Yes	Yes	Silver	7440-22-4	2		mg/kg
Yes	Yes	Yes	Sodium	7440-23-5	50		mg/kg
Non	No	No	Strontium	7440-24-6	5		mg/kg
Yes	Yes	Yes	Thallium	7440-28-0	10		mg/kg
Con	No	No	Titanium	7440-32-6	1		mg/kg
Yes	Yes	Yes	Vanadium	7440-62-2	5		mg/kg
Yes	Yes	Yes	Zinc	7440-66-6	5		mg/kg
Con	No	No	Zirconium	7440-67-7			mg/kg

ASR: 8209

Analysis Name: Metals in Solids by ICP-AES **Analysis Request No** 1

Default Report Flag: Con Analyte that is not reported from in-house analysis and must be obtained through an out-source contract lab.

Non Analyte that is not routinely reported from in-house analysis, but can be if it is needed.

Yes Analyte that is routinely reported from in-house analysis.

Requested Report Flag: No Result will not be stored or reported.

Yes Result will be stored and reported.